



Reference Guide for SwabCap[®]

Disinfecting Cap for Needlefree Connectors



Clinically proven disinfecting technology designed to reduce the risk of bacterial contamination and improve patient outcomes.



Peel

- › Remove protective cover by pulling tab.
- › Do not remove SwabCap from white plastic holder.



Push and Twist

- › Hold needlefree connector in one hand while pushing and twisting the SwabCap in a clockwise direction onto the needlefree connector.



Protect

- › After SwabCap has been twisted into place, gently pull white plastic holder off the needlefree connector.



To Remove for Access to Valve

- › Two-handed removal must be used to make sure the needlefree connector is not removed unintentionally.
- › Grasp the needlefree connector and twist SwabCap counter clockwise away from the connector. Remove and discard.
- › Inspect the needlefree connector to verify that it's connected securely to the catheter.
- › Connector is ready for access. No further swabbing is necessary after removal.

Precautions: Use by expiration date on package. Two-handed removal must be used to make sure luer access valve is not removed unintentionally. Do not use if product is dry of alcohol. Single use only.
Note: Intended to be used and handled by clinical personnel only and care should be exercised to prevent unintended handling by children.